

Sample Tubes Requirement

Dear Customer

When using our Sanger Sequencing Service (Economy Run), please consider the following requirements on the sample tubes:

Requirements of Sample Tubes



Please make sure that your samples are placed in 1.5 ml tubes. Screw cap tubes (see picture #1, #2) are the most robust and save tubes (no accidental lid opening!) If you use snap cap tubes (see picture #3) we recommend you to use Safe-Lock/Safe-Seal tubes (lower risk for accidental lid opening)!



Microsynth cannot process 2 ml, 0.5 ml and 0.2 ml sample tubes. Our highly automated process of sample preparation requires the use of 1.5 ml tubes.

Correct Labeling of 1.5 ml Sample Tubes



Please make sure that your 1.5 ml sample tubes are labeled as shown in the images on the left hand side. Important: Please fix labels tightly without any additional adhesive tape.



Microsynth's automated sequencing process is not compliant with

- Tubes where the lid is labeled (see picture #1)
- Tubes where the label is attached on the ribbed surface area below the screw (see picture #2)
- Tubes which are wrapped with parafilm (see picture #3)

Correct Packaging of Your Samples



Please make sure that your tubes are placed within a transparent plastic bag.

Need More Information?

In case you need help or more information, please do not hesitate to

- call us at +41-71-726 10 04
- or email us at sanger.support@microsynth.ch